



T&S BRASS AND BRONZE WORKS, INC.

2 Saddleback Cove / P.O. Box 1088
Travelers Rest, SC 29690

Model No.

BL-4250-02

Item No.

Travelers Rest, SC: 800-476-4103 • Simi Valley, CA: 800-423-0150 • Fax: 864-834-3518 • www.tsbrass.com



ADA Compliant

NOTES:

1. Natural Gas Pressure: 0.5 PSIG Max.
(For Natural Gas above 0.5 PSIG or Oxygen and Specialty Gases, use [BL-4010-01](#) Gas Cock.)
2. Air Pressure: 125 PSIG Max.
3. Vacuum: 30 inches of Mercury Max.
4. Certified to ANSI Z21.15a / CSA 9.1a

This Space for Architect/Engineer Approval

Job Name _____ Date _____

Model Specified _____ Quantity _____

Customer/Wholesaler _____

Contractor _____





Architect/Engineer _____



209L-AIR-NS
Indexes
Included



209L-VAC-NS
Indexes
Included



Product Specifications:
Panel Flange w/ (2) 3/8" NPT Hose Cocks

Product Compliance:
ANSI Z21.15a / CSA 9.1a
ANSI A117.1 (ADA)



T&S BRASS AND BRONZE WORKS, INC.

2 Saddleback Cove / P.O. Box 1088
Travelers Rest, SC 29690


Model No.

BL-4250-02

Item No.

Travelers Rest, SC: 800-476-4103 • Simi Valley, CA: 800-423-0150 • Fax: 864-834-3518 • www.tsbrass.com

ITEM NO.	SALES NO.	DESCRIPTION
1	209L-AIR-NS	Orange "AIR" Lab Button Index
2	209L-VAC-NS	Yellow "VAC" Lab Button Index
3	BL-4000-01	Ball Valve Hose Cock w/ GAS, AIR & VAC Indexes
8	209L-GAS-NS	Dark Blue "GAS" Lab Button Index



Product Specifications:
Panel Flange w/ (2) 3/8" NPT Hose Cocks

Product Compliance:
ANSI Z21.15a / CSA 9.1a
ANSI A117.1 (ADA)